
Midmark M11 - v1.0.0
Total Cycles: 5143

Name: RITTER 1
Sterilizer ID: V2178198
11/29/2024 12:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.5 F 0.0
2:00 153.2 F 0.1
4:00 177.0 F 0.5
6:00 197.6 F 1.6
8:00 213.8 F 5.7
10:00 226.0 F 10.1
12:00 231.9 F 8.0
14:00 240.4 F 11.8
16:00 238.3 F 10.0
18:00 237.3 F 8.8
20:00 251.2 F 15.8
22:00 263.5 F 22.8
24:00 272.2 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 271.9 F 28.7
2:00 272.5 F 29.0
3:00 272.3 F 28.7
4:00 272.0 F 28.5

Min 271.3 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.4 F 0.0
5:00 214.6 F 0.0
10:00 229.6 F 0.0
15:00 225.6 F 0.0
20:00 234.2 F 0.0
25:00 230.3 F 0.0
30:00 228.2 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 24:15
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:02:55
CYCLE COMPLETE

11/29/2024 01:20 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____