

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3279

Name: RITTER 1  
Sterilizer ID: V2178198  
05/24/2023 03:45 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 150.4 F 0.0  
2:00 149.7 F 0.0  
4:00 172.4 F 2.1  
6:00 196.1 F 6.5  
8:00 214.5 F 11.4  
10:00 224.2 F 6.3  
12:00 236.0 F 11.6  
14:00 235.1 F 8.9  
16:00 239.3 F 9.8  
18:00 247.9 F 14.0  
20:00 260.0 F 20.8  
22:00 271.1 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.3  
1:00 272.5 F 28.9  
2:00 271.4 F 28.2  
3:00 272.7 F 29.0  
4:00 271.9 F 28.4

Min 271.3 F 28.2  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.3 F 0.0  
5:00 202.9 F 0.0  
10:00 214.8 F 0.0  
15:00 217.6 F 0.0  
20:00 241.2 F 0.0  
25:00 230.2 F 0.0  
30:00 228.6 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 22:42  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:00:55  
CYCLE COMPLETE

05/24/2023 04:45 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_