
Midmark M11 - v1.0.0
Total Cycles: 2777

Name: RITTER 1
Sterilizer ID: V2178198
01/12/2023 03:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.3 F 0.0
2:00 197.0 F 0.6
4:00 214.0 F 6.1
6:00 224.0 F 11.4
8:00 229.6 F 7.6
10:00 243.2 F 13.5
12:00 242.5 F 11.9
14:00 241.5 F 10.9
16:00 255.4 F 18.1
18:00 267.5 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.5 F 29.0
2:00 271.7 F 28.2
3:00 272.1 F 28.9
4:00 272.4 F 28.8

Min 271.4 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.0 F 0.0
5:00 224.0 F 0.0
10:00 233.8 F 0.0
15:00 230.2 F 0.0
20:00 229.2 F 0.0
25:00 230.3 F 0.0
30:00 229.0 F 0.0
Drying Complete

FILLING: 1:38
HEATING: 19:24
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:17
CYCLE COMPLETE

01/12/2023 04:41 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____