

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3648

Name: RITTER 1  
Sterilizer ID: V2178198  
08/14/2023 05:20 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 205.0 F 0.0  
2:00 189.0 F 3.0  
4:00 195.3 F 7.2  
6:00 210.9 F 11.6  
8:00 221.9 F 5.3  
10:00 229.9 F 9.6  
12:00 235.8 F 10.1  
14:00 242.1 F 11.6  
16:00 239.1 F 9.9  
18:00 251.7 F 16.0  
20:00 262.3 F 22.1  
22:00 272.2 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.2  
1:00 272.0 F 28.3  
2:00 272.3 F 28.7  
3:00 271.8 F 28.7  
4:00 271.6 F 28.1

Min 271.3 F 28.1  
Max 272.5 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.2 F 0.0  
5:00 236.9 F 0.0  
10:00 222.6 F 0.0  
15:00 214.3 F 0.0  
20:00 209.2 F 0.0  
25:00 209.3 F 0.0  
30:00 208.4 F 0.0  
Drying Complete

FILLING: 1:37  
HEATING: 22:24  
STERILIZING: 4:00  
VENTING: 3:43  
DRYING: 30:00  
TOTAL CYCLE: 01:01:43  
CYCLE COMPLETE

08/14/2023 06:21 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_