
Midmark M11 - v1.0.0
Total Cycles: 5417

Name: RITTER 1
Sterilizer ID: V2178198
03/14/2025 01:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 99.1 F 0.0
2:00 101.1 F 0.9
4:00 115.2 F 2.9
6:00 139.0 F 5.6
8:00 165.8 F 8.6
10:00 200.8 F 11.9
12:00 207.2 F 3.3
14:00 215.5 F 6.8
16:00 225.1 F 10.3
18:00 235.9 F 8.8
20:00 236.6 F 10.5
22:00 236.5 F 8.4
24:00 246.5 F 13.4
26:00 257.0 F 19.0
28:00 266.3 F 25.0
30:00 271.3 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 272.3 F 29.0
1:00 272.6 F 29.4
2:00 272.0 F 29.5
3:00 271.9 F 28.9
4:00 272.5 F 29.4

Min 271.7 F 28.9
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.1 F 0.0
5:00 184.5 F 0.0
10:00 195.3 F 0.0
15:00 196.4 F 0.0
20:00 188.3 F 0.0
25:00 189.5 F 0.0
30:00 191.9 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 30:27
STERILIZING: 4:00
VENTING: 3:58
DRYING: 30:00
TOTAL CYCLE: 01:09:12
CYCLE COMPLETE

03/14/2025 02:12 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____