

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3937

Name: RITTER2  
Sterilizer ID: V2313228  
09/12/2024 01:15 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 138.1 F 0.0  
2:00 141.4 F 2.0  
4:00 154.9 F 5.2  
6:00 183.3 F 9.6  
8:00 208.9 F 8.1  
10:00 222.7 F 8.2  
12:00 235.5 F 13.4  
14:00 236.9 F 10.9  
16:00 236.1 F 9.4  
18:00 250.2 F 16.0  
20:00 261.7 F 22.7  
22:00 272.4 F 29.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.2  
1:00 272.5 F 29.6  
2:00 271.4 F 29.4  
3:00 272.4 F 29.6  
4:00 271.3 F 29.1

Min 271.3 F 29.0  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.4 F 0.0  
5:00 206.8 F 0.0  
10:00 218.6 F 0.0  
15:00 212.6 F 0.0  
20:00 208.8 F 0.0  
25:00 209.1 F 0.0  
30:00 206.3 F 0.0  
Drying Complete

FILLING: 1:06  
HEATING: 22:27  
STERILIZING: 4:00  
VENTING: 3:26  
DRYING: 30:00  
TOTAL CYCLE: 01:00:59  
CYCLE COMPLETE

09/12/2024 02:15 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_