

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2175

Name: RITTER2  
Sterilizer ID: V2313228  
03/27/2023 07:26 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 76.5 F 0.0  
2:00 82.5 F 0.0  
4:00 102.3 F 0.0  
6:00 131.0 F 0.0  
8:00 148.7 F 2.5  
10:00 174.0 F 5.7  
12:00 198.7 F 9.3  
14:00 217.6 F 13.3  
16:00 223.0 F 7.0  
18:00 234.6 F 11.3  
20:00 232.4 F 7.7  
22:00 243.2 F 13.4  
24:00 242.2 F 11.9  
26:00 254.4 F 18.2  
28:00 265.2 F 24.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.1  
1:00 272.6 F 29.7  
2:00 271.5 F 29.1  
3:00 272.6 F 29.7  
4:00 271.5 F 29.0

Min 271.4 F 28.9  
Max 272.8 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.4 F 0.0  
5:00 216.4 F 0.0  
10:00 213.5 F 0.0  
15:00 214.9 F 0.0  
20:00 213.7 F 0.0  
25:00 213.6 F 0.0  
30:00 212.4 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 29:52  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:07:50  
CYCLE COMPLETE

03/27/2023 08:33 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_