
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001481
15:48 25/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:45
PRESSURIZING 2:48
STERILIZING 3:50
136.0 C 223kPa 3:50
136.1 C 224kPa 4:20
136.3 C 226kPa 4:50
136.3 C 224kPa 5:20
136.4 C 227kPa 5:50
136.3 C 225kPa 6:20
136.3 C 225kPa 6:50
136.4 C 226kPa 7:20
136.4 C 226kPa 7:20
VENTING 7:20
AIR DRYING 8:11
CYCLE COMPLETE 9:29

Digital Signature #
BFE7FF53532EC256
