

-----  
Midmark M11 - v1.0.5  
Total Cycles: 468

Name: RITTER2  
Sterilizer ID: V2313228  
10/21/2021 10:13 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 182.7 F 0.0  
2:00 184.2 F 3.6  
4:00 206.1 F 9.6  
6:00 222.1 F 6.6  
8:00 235.4 F 11.6  
10:00 237.0 F 10.0  
12:00 237.0 F 9.3  
14:00 251.9 F 16.8  
16:00 264.4 F 24.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.9  
1:00 272.4 F 29.3  
2:00 272.2 F 29.5  
3:00 272.1 F 29.1  
4:00 272.1 F 29.5  
  
Min 271.5 F 28.7  
Max 272.9 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.0 F 0.0  
5:00 224.9 F 0.0  
10:00 231.8 F 0.0  
15:00 217.9 F 0.0  
20:00 213.1 F 0.0  
25:00 214.7 F 0.0  
30:00 213.4 F 0.0  
  
Drying Complete

FILLING: 2:31  
HEATING: 17:49  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 00:57:26  
CYCLE COMPLETE

10/21/2021 11:15 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02