
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000826
13:53 30/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:45
PRESSURIZING 5:29
STERILIZING 6:52
136.0 C 220kPa 6:52
136.4 C 222kPa 7:22
136.1 C 221kPa 7:52
136.4 C 223kPa 8:22
136.0 C 219kPa 8:52
136.4 C 223kPa 9:22
136.2 C 220kPa 9:52
136.4 C 223kPa 10:22
136.4 C 223kPa 10:22
VENTING 10:22
AIR DRYING 11:23
CYCLE COMPLETE 41:23

Digital Signature #
0F2C9EBA8B7DFEFE
