
Midmark M11 - v1.0.0
Total Cycles: 4350

Name: RITTER 1
Sterilizer ID: V2178198
03/01/2024 08:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.3 F 0.0
2:00 82.1 F 0.2
4:00 121.1 F 1.1
6:00 119.2 F 3.3
8:00 150.6 F 6.1
10:00 176.7 F 9.4
12:00 200.4 F 13.0
14:00 208.4 F 3.6
16:00 217.0 F 6.7
18:00 226.0 F 10.0
20:00 235.5 F 10.7
22:00 235.8 F 9.2
24:00 242.9 F 11.6
26:00 242.7 F 11.3
28:00 253.2 F 16.6
30:00 262.9 F 22.3
32:00 272.3 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.5 F 28.5
2:00 272.7 F 28.9
3:00 271.5 F 28.3
4:00 272.4 F 28.4

Min 271.4 F 28.1
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.6 F 0.0
5:00 226.5 F 0.0
10:00 222.2 F 0.0
15:00 226.4 F 0.0
20:00 224.1 F 0.0
25:00 191.9 F 0.0
30:00 181.4 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 33:10
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:11:24
CYCLE COMPLETE

03/01/2024 09:58 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____