
Midmark M11 - v1.0.0
Total Cycles: 2100

Name: RITTER 1
Sterilizer ID: V2178198
07/07/2022 01:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.5 F 0.0
2:00 159.9 F 1.5
4:00 182.3 F 4.6
6:00 206.4 F 9.8
8:00 215.5 F 4.9
10:00 228.1 F 8.4
12:00 238.4 F 13.1
14:00 237.2 F 10.0
16:00 239.0 F 8.5
18:00 248.4 F 14.6
20:00 260.4 F 21.2
22:00 271.1 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.4
1:00 272.3 F 29.1
2:00 272.3 F 28.7
3:00 272.2 F 28.9
4:00 272.6 F 29.0

Min 271.7 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 205.4 F 0.0
10:00 219.9 F 0.0
15:00 222.5 F 0.0
20:00 221.9 F 0.0
25:00 227.7 F 0.0
30:00 226.1 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 22:35
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:57
CYCLE COMPLETE

07/07/2022 02:23 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____