

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2884

Name: RITTER 1  
Sterilizer ID: V2178198  
02/09/2023 03:18 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 202.2 F 0.0  
2:00 197.1 F 4.7  
4:00 205.8 F 9.7  
6:00 222.4 F 6.2  
8:00 234.4 F 10.7  
10:00 234.6 F 8.0  
12:00 240.5 F 10.6  
14:00 246.9 F 13.8  
16:00 259.7 F 21.0  
18:00 271.4 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.6  
1:00 272.2 F 29.3  
2:00 272.0 F 28.8  
3:00 271.5 F 28.8  
4:00 272.7 F 29.3  
  
Min 271.3 F 28.4  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.2 F 0.0  
5:00 224.5 F 0.0  
10:00 233.7 F 0.0  
15:00 229.2 F 0.0  
20:00 226.1 F 0.0  
25:00 228.2 F 0.0  
30:00 226.6 F 0.0  
Drying Complete

FILLING: 1:59  
HEATING: 18:31  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:57:45  
CYCLE COMPLETE

02/09/2023 04:15 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_