

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1377

Name: RITTER2  
Sterilizer ID: V2313228  
08/08/2022 02:39 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 193.6 F 0.0  
2:00 187.5 F 0.6  
4:00 202.5 F 6.0  
6:00 220.6 F 11.7  
8:00 229.2 F 8.6  
10:00 238.0 F 10.4  
12:00 244.1 F 12.4  
14:00 246.5 F 13.8  
16:00 259.8 F 21.2  
18:00 271.5 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 271.9 F 29.0  
2:00 272.4 F 29.4  
3:00 272.0 F 29.6  
4:00 271.8 F 28.9  
  
Min 271.5 F 28.8  
Max 272.6 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.7 F 0.0  
5:00 223.0 F 0.0  
10:00 233.3 F 0.0  
15:00 227.5 F 0.0  
20:00 221.7 F 0.0  
25:00 223.4 F 0.0  
30:00 223.8 F 0.0  
Drying Complete

FILLING: 1:28  
HEATING: 18:32  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:57:14  
CYCLE COMPLETE

08/08/2022 03:36 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_