
Midmark M11 - v1.0.0
Total Cycles: 4511

Name: RITTER 1
Sterilizer ID: V2178198
05/01/2024 02:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.4 F 0.0
2:00 179.3 F 3.3
4:00 183.8 F 7.4
6:00 204.6 F 12.5
8:00 215.6 F 6.7
10:00 236.9 F 11.9
12:00 231.7 F 9.0
14:00 238.2 F 9.8
16:00 246.9 F 13.8
18:00 259.2 F 20.5
20:00 270.0 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.3 F 29.6
2:00 272.0 F 29.0
3:00 273.0 F 29.6
4:00 271.7 F 29.0

Min 270.9 F 28.5
Max 273.1 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.2 F 0.0
5:00 222.9 F 0.0
10:00 220.5 F 0.0
15:00 212.2 F 0.0
20:00 205.9 F 0.0
25:00 201.9 F 0.0
30:00 198.4 F 0.0
Drying Complete

FILLING: 1:37
HEATING: 20:47
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 00:59:56
CYCLE COMPLETE

05/01/2024 03:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____