
Midmark M11 - v1.0.0
Total Cycles: 605

Name: MIDMARK
Sterilizer ID: V2178198
01/14/2021 01:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 102.1 F 0.0
2:00 102.9 F 0.0
4:00 111.6 F 0.2
6:00 145.7 F 2.6
8:00 181.5 F 6.1
10:00 213.3 F 10.5
12:00 217.4 F 5.0
14:00 228.6 F 10.3
16:00 230.2 F 7.7
18:00 241.9 F 13.1
20:00 241.6 F 11.5
22:00 255.0 F 18.2
24:00 266.8 F 25.3
26:00 272.6 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.1
1:00 271.9 F 29.0
2:00 272.1 F 29.0
3:00 272.6 F 29.3
4:00 272.7 F 29.5

Min 271.8 F 29.0
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.6 F 0.0
5:00 239.7 F 0.0
10:00 225.3 F 0.0
15:00 222.3 F 0.0
20:00 212.0 F 0.0
25:00 206.7 F 0.0
30:00 212.0 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 26:09
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:04:21
CYCLE COMPLETE

01/14/2021 02:54 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____