
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000916
10:54 07/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:28
PRESSURIZING 3:30
STERILIZING 4:40
136.5 C 223kPa 4:40
136.1 C 222kPa 5:10
136.3 C 222kPa 5:40
136.4 C 223kPa 6:10
136.3 C 223kPa 6:40
136.2 C 219kPa 7:10
136.4 C 225kPa 7:40
136.4 C 221kPa 8:10
136.4 C 221kPa 8:10
VENTING 8:10
AIR DRYING 9:06
CYCLE COMPLETE 9:26

Digital Signature #
88CE99A8F41B527A
