
Midmark M11 - v1.0.0
Total Cycles: 2892

Name: RITTER 1
Sterilizer ID: V2178198
02/13/2023 01:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 135.9 F 0.0
2:00 139.6 F 2.5
4:00 163.8 F 5.9
6:00 187.1 F 10.1
8:00 206.3 F 6.4
10:00 216.9 F 7.6
12:00 229.7 F 12.3
14:00 232.2 F 8.6
16:00 240.9 F 11.5
18:00 243.2 F 12.3
20:00 255.2 F 18.5
22:00 266.1 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.6 F 29.6
2:00 271.9 F 28.9
3:00 272.0 F 29.3
4:00 272.4 F 29.4

Min 271.4 F 28.7
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.5 F 0.0
5:00 216.9 F 0.0
10:00 232.3 F 0.0
15:00 234.2 F 0.0
20:00 226.2 F 0.0
25:00 231.2 F 0.0
30:00 229.9 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 23:43
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:02:02
CYCLE COMPLETE

02/13/2023 02:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____