

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4264

Name: RITTER 1  
Sterilizer ID: V2178198  
02/02/2024 01:56 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 113.4 F 0.0  
2:00 116.9 F 1.3  
4:00 135.7 F 3.7  
6:00 172.3 F 7.3  
8:00 192.8 F 11.5  
10:00 218.6 F 3.5  
12:00 219.6 F 7.7  
14:00 232.5 F 12.2  
16:00 232.7 F 8.3  
18:00 244.8 F 12.3  
20:00 243.8 F 12.1  
22:00 255.8 F 18.2  
24:00 266.5 F 24.7  
26:00 273.1 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.2 F 28.9  
2:00 272.4 F 29.2  
3:00 272.1 F 28.6  
4:00 272.7 F 29.2

Min 271.4 F 28.4  
Max 272.9 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 165.5 F 0.0  
5:00 189.9 F 0.0  
10:00 190.0 F 0.0  
15:00 188.4 F 0.0  
20:00 188.1 F 0.0  
25:00 209.8 F 0.0  
30:00 192.8 F 0.0  
Drying Complete

FILLING: 0:60  
HEATING: 26:20  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 01:04:52  
CYCLE COMPLETE

02/02/2024 03:00 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_