
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 002519
14:55 25/10/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:12
PRESSURIZING 3:14
STERILIZING 4:19
136.0 C 220kPa 4:19
136.4 C 225kPa 4:49
CYCLE ABORTED 5:13
STOP BUTTON PRESSED
NOT STERILE
END TIME 5:13

Digital Signature #
6F4FB65E71070966
