

-----  
Midmark M11 - v1.0.5  
Total Cycles: 723

Name: RITTER2  
Sterilizer ID: V2313228  
01/24/2022 12:42 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 148.8 F 0.0  
2:00 155.5 F 3.4  
4:00 181.7 F 8.0  
6:00 206.7 F 13.5  
8:00 222.2 F 8.5  
10:00 233.9 F 10.0  
12:00 241.2 F 12.6  
14:00 242.1 F 11.9  
16:00 255.9 F 19.1  
18:00 268.0 F 26.6

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 29.1  
1:00 272.6 F 29.9  
2:00 271.6 F 29.0  
3:00 272.7 F 29.8  
4:00 271.8 F 29.0  
  
Min 271.5 F 28.9  
Max 272.8 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.5 F 0.0  
5:00 221.1 F 0.0  
10:00 203.1 F 0.0  
15:00 202.9 F 0.0  
20:00 200.1 F 0.0  
25:00 224.1 F 0.0  
30:00 224.8 F 0.0  
  
Drying Complete

FILLING: 1:31  
HEATING: 19:06  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 00:57:42  
CYCLE COMPLETE

01/24/2022 01:39 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_