
Midmark M11 - v1.0.0
Total Cycles: 2608

Name: RITTER 1
Sterilizer ID: V2178198
11/28/2022 08:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.1 F 0.0
2:00 145.7 F 0.5
4:00 169.5 F 3.8
6:00 189.9 F 7.8
8:00 208.0 F 12.2
10:00 218.6 F 5.1
12:00 229.4 F 9.5
14:00 238.8 F 10.3
16:00 239.0 F 11.0
18:00 236.5 F 8.6
20:00 249.3 F 15.2
22:00 260.8 F 21.6
24:00 271.3 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.5 F 29.3
2:00 271.5 F 28.5
3:00 272.5 F 29.4
4:00 271.9 F 28.7

Min 271.4 F 28.4
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.0 F 0.0
5:00 201.3 F 0.0
10:00 206.3 F 0.0
15:00 215.5 F 0.0
20:00 217.9 F 0.0
25:00 212.4 F 0.0
30:00 210.5 F 0.0
Drying Complete

FILLING: 0:36
HEATING: 24:41
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:02:28
CYCLE COMPLETE

11/28/2022 09:16 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____