
Midmark M11 - v1.0.5
Total Cycles: 376

Name: RITTER2
Sterilizer ID: V2313228
09/20/2021 06:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.6 F 0.0
2:00 73.6 F 0.0
4:00 74.3 F 0.1
6:00 89.5 F 0.2
8:00 128.6 F 1.1
10:00 167.1 F 4.9
12:00 189.7 F 8.8
14:00 207.1 F 12.7
16:00 215.5 F 6.8
18:00 228.7 F 11.2
20:00 231.3 F 7.9
22:00 242.1 F 12.4
24:00 243.7 F 12.6
26:00 256.0 F 19.0
28:00 267.1 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.9
1:00 272.4 F 29.6
2:00 271.2 F 28.9
3:00 272.4 F 29.6
4:00 271.2 F 28.8

Min 271.0 F 28.8
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.2 F 0.0
5:00 223.9 F 0.0
10:00 225.5 F 0.0
15:00 230.1 F 0.0
20:00 221.7 F 0.0
25:00 209.5 F 0.0
30:00 197.2 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 29:27
STERILIZING: 4:00
VENTING: 2:54
DRYING: 30:00
TOTAL CYCLE: 01:07:16
CYCLE COMPLETE

09/20/2021 07:45 AM
ACCEPTED: X
REJECTED:
Operator: U01