
Midmark M11 - v1.0.5
Total Cycles: 4044

Name: RITTER2
Sterilizer ID: V2313228
10/21/2024 10:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.7 F 0.0
2:00 70.2 F 0.0
4:00 92.9 F 0.0
6:00 119.0 F 0.1
8:00 146.4 F 0.2
10:00 169.5 F 0.6
12:00 194.9 F 3.2
14:00 205.3 F 6.5
16:00 214.9 F 9.8
18:00 225.0 F 11.2
20:00 230.4 F 8.7
22:00 240.5 F 13.3
24:00 239.8 F 10.8
26:00 236.9 F 9.2
28:00 250.4 F 15.6
30:00 261.4 F 21.9
32:00 271.7 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.6 F 29.0
2:00 271.6 F 28.7
3:00 272.4 F 29.0
4:00 271.6 F 28.7

Min 271.5 F 28.3
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.2 F 0.0
5:00 176.9 F 0.0
10:00 206.0 F 0.0
15:00 203.8 F 0.0
20:00 201.5 F 0.0
25:00 201.9 F 0.0
30:00 206.8 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 32:35
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:10:44
CYCLE COMPLETE

10/21/2024 11:24 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____