

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 003341  
11:13 30/03/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:49  
PRESSURIZING 1:55  
STERILIZING 2:55  
136.4 C 225kPa 2:55  
136.3 C 223kPa 3:25  
136.4 C 223kPa 3:55  
136.4 C 223kPa 4:25  
136.2 C 223kPa 4:55  
136.3 C 223kPa 5:25  
136.3 C 224kPa 5:55  
136.3 C 223kPa 6:25  
136.3 C 223kPa 6:25  
VENTING 6:25  
AIR DRYING 7:18  
CYCLE COMPLETE 20:26

-----  
Digital Signature #  
16CC8978A6AF8EDA  
-----