
Midmark M11 - v1.0.5
Total Cycles: 3094

Name: RITTER2
Sterilizer ID: V2313228
11/07/2023 01:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.4 F 0.0
2:00 148.5 F 1.5
4:00 163.2 F 4.7
6:00 187.5 F 8.9
8:00 208.1 F 12.0
10:00 227.1 F 7.7
12:00 235.5 F 12.6
14:00 236.1 F 9.6
16:00 237.4 F 9.5
18:00 247.4 F 14.3
20:00 259.2 F 20.9
22:00 270.1 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.9
1:00 272.4 F 29.4
2:00 271.1 F 29.1
3:00 272.4 F 29.5
4:00 271.1 F 28.9

Min 271.0 F 28.8
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.8 F 0.0
5:00 234.4 F 0.0
10:00 226.2 F 0.0
15:00 223.4 F 0.0
20:00 220.5 F 0.0
25:00 220.5 F 0.0
30:00 220.5 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 22:54
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:01:22
CYCLE COMPLETE

11/07/2023 02:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____