
Midmark M11 - v1.0.5
Total Cycles: 630

Name: RITTER2
Sterilizer ID: V2313228
12/14/2021 09:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 118.4 F 0.0
2:00 121.2 F 1.0
4:00 136.6 F 3.2
6:00 162.2 F 6.6
8:00 186.0 F 10.9
10:00 205.7 F 3.9
12:00 218.3 F 7.7
14:00 230.1 F 12.4
16:00 233.0 F 9.1
18:00 239.5 F 10.5
20:00 244.7 F 13.2
22:00 256.3 F 19.4
24:00 266.9 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.2
1:00 272.9 F 29.9
2:00 271.8 F 29.1
3:00 272.1 F 29.7
4:00 272.8 F 29.7

Min 271.5 F 29.0
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.4 F 0.0
5:00 195.9 F 0.0
10:00 198.6 F 0.0
15:00 201.4 F 0.0
20:00 197.2 F 0.0
25:00 197.4 F 0.0
30:00 202.9 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 25:32
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:03:45
CYCLE COMPLETE

12/14/2021 10:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____