
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000048
9:43 08/09/2020
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:34
PRESSURIZING 7:54
STERILIZING 9:51
136.0 C 223kPa 9:51
136.4 C 227kPa 10:22
136.2 C 225kPa 10:52
136.5 C 227kPa 11:22
136.5 C 226kPa 11:52
136.3 C 226kPa 12:22
136.4 C 225kPa 12:52
136.4 C 227kPa 13:21
VENTING 13:22
AIR DRYING 14:39
CYCLE COMPLETE 44:39

Digital Signature #
55ACB493E0A79482
