

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3714

Name: RITTER2  
Sterilizer ID: V2313228  
06/18/2024 05:40 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 103.3 F 0.0  
2:00 105.3 F 0.5  
4:00 132.2 F 2.7  
6:00 158.6 F 5.9  
8:00 184.9 F 10.0  
10:00 207.6 F 7.2  
12:00 219.5 F 7.5  
14:00 231.4 F 12.0  
16:00 233.6 F 8.5  
18:00 241.7 F 11.6  
20:00 244.2 F 12.6  
22:00 256.4 F 19.0  
24:00 267.4 F 25.9

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 28.8  
1:00 272.2 F 29.6  
2:00 271.7 F 28.8  
3:00 272.2 F 29.6  
4:00 271.9 F 28.9

Min 271.1 F 28.6  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.7 F 0.0  
5:00 223.9 F 0.0  
10:00 195.5 F 0.0  
15:00 187.9 F 0.0  
20:00 186.8 F 0.0  
25:00 187.4 F 0.0  
30:00 186.8 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 25:16  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:03:33  
CYCLE COMPLETE

06/18/2024 06:43 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_