
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001597
15:54 11/18/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:38
PRESSURIZING 6:14
STERILIZING 7:43
136.0 C 220kPa 7:43
136.2 C 221kPa 8:13
136.4 C 223kPa 8:43
136.3 C 223kPa 9:13
136.0 C 220kPa 9:43
136.2 C 219kPa 10:13
136.2 C 221kPa 10:43
136.2 C 222kPa 11:13
VENTING 11:13
AIR DRYING 12:13
CYCLE COMPLETE 19:12

Digital Signature #
8D5F8B9D739C9602
