

-----  
Midmark M11 - v1.0.5  
Total Cycles: 845

Name: RITTER2  
Sterilizer ID: V2313228  
02/28/2022 01:58 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 148.4 F 0.0  
2:00 149.9 F 3.0  
4:00 170.4 F 6.9  
6:00 194.4 F 11.9  
8:00 213.6 F 6.0  
10:00 227.9 F 11.3  
12:00 232.1 F 8.7  
14:00 239.1 F 10.3  
16:00 247.0 F 14.2  
18:00 259.7 F 21.3  
20:00 271.2 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.0  
1:00 272.4 F 29.8  
2:00 272.0 F 29.1  
3:00 271.9 F 29.6  
4:00 272.6 F 29.5

Min 271.4 F 28.9  
Max 272.9 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.9 F 0.0  
5:00 212.3 F 0.0  
10:00 217.5 F 0.0  
15:00 210.2 F 0.0  
20:00 216.5 F 0.0  
25:00 221.8 F 0.0  
30:00 219.3 F 0.0  
Drying Complete

FILLING: 1:18  
HEATING: 20:32  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:58:58  
CYCLE COMPLETE

02/28/2022 02:56 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_