

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002045  
11:06 16/09/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:05  
PRESSURIZING 2:10  
STERILIZING 3:09  
136.4 C 225kPa 3:09  
136.4 C 227kPa 3:39  
136.4 C 226kPa 4:09  
136.4 C 227kPa 4:39  
136.4 C 226kPa 5:09  
136.4 C 226kPa 5:39  
136.4 C 223kPa 6:09  
136.3 C 225kPa 6:39  
136.3 C 225kPa 6:39  
VENTING 6:39  
AIR DRYING 7:32  
CYCLE COMPLETE 7:46  
-----

Digital Signature #  
4D8E7B461079F864  
-----