
Midmark M11 - v1.0.0
Total Cycles: 1492

Name: RITTER 1
Sterilizer ID: V2178198
12/30/2021 07:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.1 F 0.0
2:00 74.6 F 0.0
4:00 97.5 F 0.0
6:00 130.9 F 0.1
8:00 155.9 F 0.4
10:00 182.6 F 2.6
12:00 206.9 F 4.8
14:00 217.8 F 7.4
16:00 226.8 F 10.6
18:00 234.3 F 8.3
20:00 233.9 F 9.4
22:00 242.4 F 11.4
24:00 242.1 F 11.3
26:00 237.8 F 9.2
28:00 251.3 F 15.7
30:00 262.5 F 22.1
32:00 272.7 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.0
1:00 272.5 F 28.6
2:00 271.3 F 28.1
3:00 272.6 F 28.7
4:00 271.4 F 27.9

Min 271.3 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.6 F 0.0
5:00 225.3 F 0.0
10:00 210.2 F 0.0
15:00 236.0 F 0.0
20:00 225.5 F 0.0
25:00 228.6 F 0.0
30:00 211.7 F 0.0

Drying Complete

FILLING: 0:53
HEATING: 32:20
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:10:22
CYCLE COMPLETE

12/30/2021 10:06 AM
ACCEPTED: X
REJECTED:
Operator: U02