

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001418  
12:11 09/09/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:11  
PRESSURIZING 4:43  
STERILIZING 5:59  
136.0 C 220kPa 5:59  
135.8 C 219kPa 6:29  
136.4 C 223kPa 6:59  
136.4 C 224kPa 7:29  
136.0 C 217kPa 7:59  
136.4 C 223kPa 8:29  
136.4 C 223kPa 8:59  
136.4 C 221kPa 9:29  
VENTING 9:29  
AIR DRYING 10:30  
CYCLE COMPLETE 10:32

-----  
Digital Signature #  
546C5FAAB5C45ACB  
-----