
Midmark M11 - v1.0.0
Total Cycles: 3683

Name: RITTER 1
Sterilizer ID: V2178198
08/25/2023 10:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.0 F 0.0
2:00 184.7 F 2.1
4:00 179.7 F 5.5
6:00 208.9 F 10.1
8:00 215.8 F 5.8
10:00 228.9 F 8.3
12:00 240.9 F 13.1
14:00 238.3 F 10.0
16:00 237.3 F 8.8
18:00 248.4 F 14.3
20:00 259.6 F 20.5
22:00 269.9 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.2 F 28.4
2:00 272.6 F 28.9
3:00 271.6 F 28.1
4:00 272.6 F 28.7

Min 271.5 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.6 F 0.0
5:00 236.8 F 0.0
10:00 229.4 F 0.0
15:00 232.3 F 0.0
20:00 226.3 F 0.0
25:00 229.2 F 0.0
30:00 230.7 F 0.0
Drying Complete

FILLING: 1:29
HEATING: 22:54
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:01:49
CYCLE COMPLETE

08/25/2023 11:22 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____