
Midmark M11 - v1.0.5
Total Cycles: 2365

Name: RITTER2
Sterilizer ID: V2313228
05/15/2023 10:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.7 F 0.0
2:00 184.2 F 3.2
4:00 191.6 F 7.6
6:00 210.3 F 13.2
8:00 224.5 F 8.3
10:00 236.3 F 12.0
12:00 240.2 F 11.6
14:00 239.2 F 10.4
16:00 252.6 F 17.1
18:00 264.0 F 23.8
20:00 270.9 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.1
1:00 272.3 F 29.3
2:00 272.9 F 29.7
3:00 271.7 F 29.2
4:00 272.4 F 29.4

Min 271.7 F 29.0
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.8 F 0.0
5:00 208.8 F 0.0
10:00 219.9 F 0.0
15:00 224.6 F 0.0
20:00 225.0 F 0.0
25:00 226.4 F 0.0
30:00 224.3 F 0.0
Drying Complete

FILLING: 1:51
HEATING: 20:41
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:44
CYCLE COMPLETE

05/15/2023 11:20 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____