
Midmark M11 - v1.0.5
Total Cycles: 1371

Name: RITTER2
Sterilizer ID: V2313228
08/04/2022 02:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.5 F 0.0
2:00 189.0 F 0.7
4:00 191.9 F 5.3
6:00 211.4 F 10.3
8:00 224.7 F 6.9
10:00 236.6 F 12.2
12:00 237.2 F 10.2
14:00 235.8 F 8.8
16:00 250.1 F 15.9
18:00 262.1 F 22.8
20:00 272.7 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.0
1:00 271.8 F 29.5
2:00 272.4 F 29.4
3:00 271.5 F 28.8
4:00 272.0 F 29.7

Min 271.3 F 28.8
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.1 F 0.0
5:00 218.0 F 0.0
10:00 239.2 F 0.0
15:00 237.6 F 0.0
20:00 227.8 F 0.0
25:00 223.8 F 0.0
30:00 221.2 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 20:13
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:60
CYCLE COMPLETE

08/04/2022 03:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____