
Midmark M11 - v1.0.0
Total Cycles: 3694

Name: RITTER 1
Sterilizer ID: V2178198
08/28/2023 05:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.1 F 0.0
2:00 167.3 F 3.7
4:00 181.8 F 7.8
6:00 210.3 F 12.5
8:00 220.6 F 6.2
10:00 231.9 F 11.0
12:00 230.7 F 7.4
14:00 244.6 F 13.5
16:00 243.5 F 12.2
18:00 256.5 F 18.8
20:00 267.7 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.1 F 28.6
2:00 272.3 F 29.1
3:00 271.3 F 28.2
4:00 272.2 F 29.0

Min 271.2 F 28.2
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.9 F 0.0
5:00 220.5 F 0.0
10:00 233.5 F 0.0
15:00 231.2 F 0.0
20:00 227.3 F 0.0
25:00 223.4 F 0.0
30:00 224.8 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 21:18
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:00:06
CYCLE COMPLETE

08/28/2023 06:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____