
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007563
8:37 18/06/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:40
PRESSURIZING 3:41
STERILIZING 4:50
136.5 C 224kPa 4:50
136.4 C 224kPa 5:20
136.4 C 221kPa 5:50
136.4 C 223kPa 6:20
136.3 C 223kPa 6:50
136.3 C 220kPa 7:20
136.4 C 223kPa 7:50
136.4 C 224kPa 8:20
136.4 C 224kPa 8:20
VENTING 8:20
AIR DRYING 9:17
CYCLE COMPLETE 39:17

Digital Signature #
4523099053C83777
