
Midmark M11 - v1.0.5
Total Cycles: 1510

Name: RITTER2
Sterilizer ID: V2313228
09/20/2022 02:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 206.0 F 0.0
2:00 194.5 F 3.8
4:00 202.5 F 9.5
6:00 219.5 F 6.6
8:00 235.1 F 11.0
10:00 235.5 F 9.1
12:00 239.4 F 9.7
14:00 249.5 F 15.4
16:00 262.0 F 22.5
18:00 272.7 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.6 F 29.5
2:00 271.5 F 28.8
3:00 272.7 F 29.7
4:00 272.1 F 29.1

Min 271.5 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.3 F 0.0
5:00 223.5 F 0.0
10:00 231.2 F 0.0
15:00 230.2 F 0.0
20:00 227.4 F 0.0
25:00 229.4 F 0.0
30:00 225.7 F 0.0
Drying Complete

FILLING: 3:07
HEATING: 18:21
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:40
CYCLE COMPLETE

09/20/2022 03:12 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____