

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 007282  
9:26 29/04/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:34  
PRESSURIZING 2:38  
STERILIZING 3:35  
136.0 C 220kPa 3:35  
136.2 C 223kPa 4:05  
136.4 C 221kPa 4:35  
136.2 C 223kPa 5:05  
136.3 C 224kPa 5:35  
136.1 C 222kPa 6:05  
136.4 C 224kPa 6:35  
136.3 C 224kPa 7:05  
136.3 C 224kPa 7:05  
VENTING 7:05  
AIR DRYING 7:55  
CYCLE COMPLETE 8:04

-----  
Digital Signature #  
7A06CAA6A6991045  
-----