
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000540
9:33 09/07/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 2:38
STERILIZING 3:37
136.0 C 220kPa 3:37
136.4 C 233kPa 4:07
136.3 C 230kPa 4:37
136.4 C 231kPa 5:07
136.0 C 228kPa 5:37
136.4 C 231kPa 6:07
136.3 C 231kPa 6:37
136.4 C 232kPa 7:07
136.4 C 232kPa 7:07
VENTING 7:07
AIR DRYING 7:55
CYCLE COMPLETE 22:55

Digital Signature #
1258D4EA7F8970B5
