
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002616
13:32 26/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:56
PRESSURIZING 2:00
STERILIZING 2:56
136.4 C 223kPa 2:56
136.3 C 224kPa 3:27
136.1 C 222kPa 3:57
136.1 C 223kPa 4:27
136.2 C 223kPa 4:57
136.0 C 221kPa 5:27
136.2 C 224kPa 5:57
136.2 C 224kPa 6:27
136.2 C 224kPa 6:27
VENTING 6:27
AIR DRYING 7:16
CYCLE COMPLETE 37:16

Digital Signature #
340B137E6AC7C817
