

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3898

Name: RITTER 1  
Sterilizer ID: V2178198  
10/13/2023 08:11 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.4 F 0.0  
2:00 76.1 F 0.0  
4:00 127.7 F 0.0  
6:00 158.7 F 0.1  
8:00 174.7 F 0.3  
10:00 200.9 F 3.5  
12:00 215.2 F 6.7  
14:00 224.2 F 10.0  
16:00 233.0 F 13.5  
18:00 225.8 F 6.9  
20:00 236.1 F 10.9  
22:00 233.0 F 7.4  
24:00 243.4 F 12.1  
26:00 239.5 F 10.0  
28:00 252.2 F 16.2  
30:00 262.7 F 22.4  
32:00 272.7 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 272.3 F 28.4  
1:00 272.7 F 28.8  
2:00 272.4 F 29.1  
3:00 271.6 F 28.4  
4:00 272.6 F 28.6  
  
Min 271.5 F 28.3  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 176.5 F 0.0  
5:00 216.9 F 0.0  
10:00 237.6 F 0.0  
15:00 229.6 F 0.0  
20:00 215.7 F 0.0  
25:00 213.7 F 0.0  
30:00 210.2 F 0.0  
  
Drying Complete

FILLING: 0:49  
HEATING: 32:57  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:11:06  
CYCLE COMPLETE

10/13/2023 09:22 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_