
Midmark M11 - v1.0.0
Total Cycles: 4165

Name: RITTER 1
Sterilizer ID: V2178198
12/28/2023 02:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 130.8 F 0.0
2:00 131.9 F 1.9
4:00 147.8 F 4.9
6:00 208.7 F 9.0
8:00 208.8 F 8.5
10:00 218.3 F 6.4
12:00 230.0 F 10.9
14:00 229.5 F 6.7
16:00 243.1 F 12.3
18:00 239.8 F 10.3
20:00 252.2 F 16.4
22:00 263.3 F 22.8
24:00 271.8 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.0 F 28.4
2:00 272.4 F 28.7
3:00 272.2 F 29.0
4:00 271.4 F 28.2

Min 271.2 F 28.2
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.0 F 0.0
5:00 243.1 F 0.0
10:00 198.4 F 0.0
15:00 211.6 F 0.0
20:00 179.6 F 0.0
25:00 174.5 F 0.0
30:00 174.1 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 24:09
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:02:37
CYCLE COMPLETE

12/28/2023 03:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____