

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 002496  
9:40 10/02/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:20  
PRESSURIZING 2:24  
STERILIZING 3:17  
136.0 C 220kPa 3:17  
136.3 C 225kPa 3:47  
136.4 C 225kPa 4:17  
136.2 C 224kPa 4:47  
136.1 C 223kPa 5:17  
136.3 C 223kPa 5:47  
136.4 C 225kPa 6:17  
136.4 C 225kPa 6:47  
136.4 C 225kPa 6:47  
VENTING 6:47  
AIR DRYING 7:43  
CYCLE COMPLETE 12:09

-----  
Digital Signature #  
3050B48D44EDC623  
-----