
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003849
10:27 21/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:29
PRESSURIZING 1:35
STERILIZING 2:43
136.0 C 220kPa 2:43
136.1 C 223kPa 3:13
136.0 C 223kPa 3:43
136.3 C 224kPa 4:13
136.4 C 224kPa 4:43
136.3 C 224kPa 5:13
136.2 C 223kPa 5:43
136.4 C 224kPa 6:13
136.4 C 224kPa 6:13
VENTING 6:14
AIR DRYING 7:07
CYCLE COMPLETE 7:42

Digital Signature #
44B740EE3FDF3F24
