

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2416

Name: RITTER 1  
Sterilizer ID: V2178198  
10/12/2022 07:32 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.3 F 0.0  
2:00 75.8 F 0.0  
4:00 86.1 F 0.0  
6:00 122.3 F 0.2  
8:00 145.6 F 2.8  
10:00 168.1 F 6.1  
12:00 187.2 F 9.8  
14:00 205.0 F 13.5  
16:00 225.0 F 5.5  
18:00 230.8 F 9.3  
20:00 237.8 F 13.5  
22:00 238.3 F 9.9  
24:00 237.5 F 8.9  
26:00 247.3 F 13.5  
28:00 258.2 F 19.5  
30:00 268.5 F 25.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.4  
1:00 272.0 F 28.3  
2:00 272.0 F 28.5  
3:00 272.2 F 28.7  
4:00 272.7 F 28.9

Min 271.7 F 28.3  
Max 272.9 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.3 F 0.0  
5:00 228.4 F 0.0  
10:00 201.9 F 0.0  
15:00 213.1 F 0.0  
20:00 203.2 F 0.0  
25:00 206.0 F 0.0  
30:00 191.4 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 31:48  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:09:47  
CYCLE COMPLETE

10/12/2022 08:41 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_