
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 005526
9:47 15/05/2023

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:57
PRESSURIZING 2:00
STERILIZING 2:59
136.4 C 221kPa 2:59
136.4 C 225kPa 3:29
136.5 C 223kPa 3:59
136.3 C 222kPa 4:29
136.4 C 224kPa 4:59
136.4 C 222kPa 5:29
136.2 C 220kPa 5:59
136.1 C 223kPa 6:29
136.1 C 223kPa 6:29
VENTING 6:29
AIR DRYING 7:19
CYCLE COMPLETE 11:49

Digital Signature #
D09BDF60CA4C6703
