
Midmark M11 - v1.0.0
Total Cycles: 844

Name: RITTER 1
Sterilizer ID: V2178198
05/27/2021 11:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.0 F 0.0
2:00 182.1 F 3.0
4:00 196.2 F 7.4
6:00 211.5 F 12.1
8:00 220.1 F 6.2
10:00 234.1 F 11.3
12:00 235.1 F 8.6
14:00 244.3 F 11.2
16:00 246.4 F 13.4
18:00 257.8 F 19.6
20:00 268.1 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.4
1:00 272.4 F 29.0
2:00 272.5 F 28.8
3:00 271.7 F 28.3
4:00 272.8 F 29.2

Min 271.6 F 28.3
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.2 F 0.0
5:00 236.0 F 0.0
10:00 227.8 F 0.0
15:00 230.5 F 0.0
20:00 229.6 F 0.0
25:00 229.1 F 0.0
30:00 223.3 F 0.0
Drying Complete

FILLING: 1:53
HEATING: 21:15
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:00:46
CYCLE COMPLETE

05/27/2021 12:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____